



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/758,777	01/16/2004	James A. Baker	9222.17777-CON DIV	6325
26308	7590	11/15/2005	EXAMINER	
RYAN KROMHOLZ & MANION, S.C. POST OFFICE BOX 26618 MILWAUKEE, WI 53226			PEFFLEY, MICHAEL F	
			ART UNIT	PAPER NUMBER

3739

DATE MAILED: 11/15/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

**Office Action Summary**

Application No.

10/758,777

Applicant(s)

BAKER, JAMES A.

Examiner

Michael Peffley

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 16 January 2004.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-5 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-5 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☒ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 16 January 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☒ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date 1/16/04.
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_.
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: \_\_\_\_\_.

***Specification***

The disclosure is objected to because of the following informalities: the "Reference to Related Applications" section should be amended to provide the most current status (i.e. US Patent Number) of the related applications.

Appropriate correction is required.

***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1, 4 and 5 are rejected under 35 U.S.C. 102(e) as anticipated by Swanson et al (6,001,093).

Swanson et al disclose a catheter and method for treating tissue that comprises a contacting first and second tissue regions with first and second electrodes connected to an RF source, and contacting a third tissue region with a third electrode that is not connected to the RF source. In particular, Figure 33 shows the device that includes a plurality of "active" electrodes (A) and inactive electrodes (N) on the catheter device, and column 24 discloses the depicted embodiment. The inactive electrodes are provided in order to create interrupted ablation lesions (Figures 34 and 35). However, in as much as the inactive electrodes are in proximity to the active electrodes in much

Art Unit: 3739

the same way as applicant's "channeling" electrode is merely an electrode disposed between two active electrodes, the examiner maintains that RF current would inherently pass through the inactive electrodes, even if there is insufficient current to create a lesion in the location of the inactive electrodes. Swanson et al disclose the particular duration of the treatments at column 24, lines 15-20.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2 and 3 are rejected under 35 U.S.C. 103(a) as being unpatentable over Swanson et al ('093).

The Swanson et al device has been addressed above. Swanson et al disclose that the device delivers about 30 watts of power in column 24, lines 15-20. The examiner maintains that the exact power level provided would be an obvious determination for the skilled artisan and would depend on the specific tissue and procedure being performed. As such, the examiner maintains that one of ordinary skill in the art would obvious consider slightly lower power ranges (i.e. 10-25 watts) as deemed appropriate for the particular procedure.

***Double Patenting***

The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent

Art Unit: 3739

and to prevent possible harassment by multiple assignees. A nonstatutory obviousness-type double patenting rejection is appropriate where the conflicting claims are not identical, but at least one examined application claim is not patentably distinct from the reference claim(s) because the examined application claim is either anticipated by, or would have been obvious over, the reference claim(s). See, e.g., *In re Berg*, 140 F.3d 1428, 46 USPQ2d 1226 (Fed. Cir. 1998); *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) or 1.321(d) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent either is shown to be commonly owned with this application, or claims an invention made as a result of activities undertaken within the scope of a joint research agreement.

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

Claims 1-5 are rejected on the ground of nonstatutory obviousness-type double patenting as being unpatentable over claims 12-18 of U.S. Patent No. 6,169,926.

Although the conflicting claims are not identical, they are not patentably distinct from each other because the '926 patented claims are more narrow and recite the bipolar application of RF energy, which limitation is deemed to be an obvious modification.


### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael Peffley whose telephone number is (571) 272-4770. The examiner can normally be reached on Mon-Fri from 6am-3pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

  
Michael Peffley  
Primary Examiner  
Art Unit 3739

mp  
November 14, 2005